CERTIFICATION

LANGER
EMV-Technik

Nöthnitzer Hang 31
Germany 01728 Bannewitz
Phone: +49 (351) 43 00 93 – 0
Fax: +49 (351) 43 00 93 – 22
email: mail@langer-emv.de

CERTIFICATE OF CALIBRATION

Cert. No. 090210QSHH150

Muster

This Probe has been individually calibrated using the following procedure for monitoring the frequency response:

Publication:
Magnetic Near Field Probe

Calibration
Procedure: LACPICR01
Uncertainty: +/- 1.5 dB

Date of Cal.: Feb 02, 2012
Cal. Interval: 12 months

Model: ICR HH150
Serial: #999#999

Manufacturer:
Langer EMV–Technik

Temperature: 20°C
Humidity: 51 %



Test and Measurement Equipment

Manufacturer	Model	Serial	Due Date
Rohde & Schwarz	FLS-6	100407	2010/03/30

TG = 100dBµV, Att = 5dB, RBW = 1MHz, VBW = 100kHz, SWT = 120ms, normalized

C. Stange

Dipl.-Ing. C. Stange
Development

LANGER
EMV-Technik GmbH
Nöthnitzer Hang 31
DE-01728 Bannewitz

This calibration is in compliance with the International EMC Standard of ICs IEC 61967. Supporting Documentation relative to traceability is on file and available for examination upon request.

This certificate shall not be reproduced except in full without the written approval of Langer EMV-Technik.

LANGER EMV-Technik GmbH / www.langer-emv.com

IC Test System - IC-Scanner_ICR Sonden 03/13